

## CERTIFICATE No. 20/2676/00/0161

Has obtained EU TYPE EXAMINATION in compliance with what is stated in Regulation (EU) 2016/425 and in agreement with the applicable test procedures and technical specifications

Respiratory protection against COVID-19 only, according to the following standard/s:

 PPE-R / 02.075 version 1. Filter half mask to protect against COVID-19 only on the basis of Commission Recommendation (EU) 2020/403 of March 13, 2020 on conformity assessment and procedures for Market surveillance in the context of the threat posed by COVID-19.

Having achieved the performance requirements specified in Technical Test Report No. 2020EC0393UE and the PPE s Technical Documentation.

Description of the PPE:

-The filtering half mask is composed by a main body with two elastic ear bands and nose clip.

2020EC0393UE.

It shall be the manufacturer's responsibility to provide specific information of this certificate and the tested levels of protection.

The CAT. III PPE shall only be used in conjunction with one of the conformity assessment procedures according to module C2 or module D described in article 19 letter c) of the Regulation (EU) 2016/425.

Digitally Signed by: Silvia Devesa

Date: 03/06/2020 12:19:26 Liberthon: Alloy

Silvia Devesa Valencia Laboratory Subdirector and Innovation

Date of issue of the Certificate: 1st of June 2020 Date of expiry: 1st of June 2025